



Test Results | About Us | Education | Links | Contact Us

get your

Print Date: 1/4/2012

HOLY MACKEREL LLC  
196 FORBES AVE.  
ST. PAUL, MN 55102

Test # QK9797  
Customer # BK575

Phone: 651-222-4659 Operator: Date Tested: 12/20/2011  
Brand of Sterilizer: Tuttnauer Model: Serial No: 2004971/A  
Sterilizing Agent: Steam Time: Temp: Pressure:

Receive Date:	12:21:03 PM	Strips: 2	Culture Readings		
Transfer Date:	12:21:03 PM		Strip	Result	Date
Test Lot #:	1116804	Exp. Date: 5/31/2013	Test Strip #1	Negative	
Media Lot #:	11361B	Exp. Date: 12/27/2012	Test Strip #2	Negative	
Incubation Temp.:	55° C		Control Strip	Positive	
Final Review Date:	4:31:58 PM				



Service Performed through:

**Results:**

Passed. No growth observed after incubation period.

Retain this completed form as part of your records as verification that a Biological Monitoring Service is utilized as a part of a formal Asepsis Program.

Luda Chalagina

Technician, Microbiological Group

Director, Microbiological Group

Test Results | About BMS | Education | Links | Contact Us

© 2001 Biological Monitoring Systems Inc. | All rights reserved. | Last Modified 3.28.2002



[Test Results](#) | 
 [About Us](#) | 
 [Education](#) | 
 [Links](#) | 
 [Contact Us](#)

# get your Test Results

from biological monitoring service

Print Date: 1/4/2012

HOLY MACKEREL LLC  
 196 FORBES AVE.  
 ST. PAUL, MN 55102

Test # QK9823  
 Customer # BK575

Phone: 651-222-4659      Operator:      Date Tested: 12/20/2011  
 Brand of Sterilizer: SciCan      Model: Statim      Serial No: 232876AA  
 Sterilizing Agent: Steam      Time:      Temp:      Pressure:

Receive Date:	1/3/2012 12:35:03 PM	Strips: 2	Culture Readings		
Transfer Date:	1/3/2012 12:35:03 PM		Strip	Result	Date
Test Lot #:	1116804	Exp. Date: 5/31/2013	Test Strip #1	Negative	
Media Lot #:	11361B	Exp. Date: 12/27/2012	Test Strip #2	Negative	
Incubation Temp.:	55° C		Control Strip	Positive	
Final Review Date:	1/4/2012 4:31:58 PM				



Service Performed through:

**Results:**

Passed. No growth observed after incubation period.

Retain this completed form as part of your records as verification that a Biological Monitoring Service is utilized as a part of a formal Asepsis Program.

Luda Chalagina

Technician, Microbiological Group

Director, Microbiological Group

[Test Results](#) | 
 [About BMS](#) | 
 [Education](#) | 
 [Links](#) | 
 [Contact Us](#)

© 2001 Biological Monitoring Systems Inc. | All rights reserved. | Last Modified 3.28.2002